

TRIGLYCERIDES

Enzymatic-spectrophotometric
GLYCEROL PHOSPHATE OXIDASE/PEROXIDASE

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **AU 5200**

Reagent preparation

Reagent is ready to be used.

Instrument settings

Operation	YES	Normal Value	H	150
Sample Volume	3 µL		L	60
Sample Volume Repeat Run	2 µL	Dynamic Range	H	600
Reagent Volume R1	250 µL		L	0
R2	0 µL	Panic Value	H	600
W3 Parameters Prime (Bottle to Valve) R1	YES		L	0
R2		Reagent OD Range		
W3 Parameters Prime (Bottle to Valve) R1	100	First Point	H	1.400
R2			L	-0.200
Method	END	Last Point	H	1.400
Wavelength 1	520		L	-0.200
2	600	Decision Level for Repeat Run	H	600
Reaction Slope	+		L	0
Measuring Point S	0	R. Dispense by Black Rack		1
E	9	R. Dispense by Sample Interval		0
OD Value Range Max		W6 Bottle Side Cleaning		Yes
Min				
% Variance				
Calibration Type	AB			
ODO Conc.	0 mg/dL			
Conc.	*	* Enter the assigned standard value		