ALKALINE PHOSPHATASE (ALP)

Continuous-spectrophotometric
DEA BUFFER

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **AUTOLAB (=CORMAY LIVIA, SABA)**

Reagent preparation

Working Reagent: Pour the contents of a Reagent B bottle into a Reagent A bottle. Mix gently.
Stable for 2 months at 2-8ºC.

Instrument settings

Test	ALP DEA	Standard value	*
Description	ALP DEA	Standard position	...
Unit	U/L	Factor	XXX
Normal value minimum	90	New factor memorization	0
maximum	280	Dilution ratio: 1:	2
Nº of decimals in result	0	Incubation time (sec)	30
Model	0	Time for Reagent 2	-
Type	2	T1	12
Nº of blanks	1	T2	50
Blank value minimum	0.150	Number of aspiration washes	2
maximum	0.800	Number of sampling washes	2
Blank correction	0	Volume of wash solution	500
Linearity	690	Prediluted sample group	1
Filter 1	405	Report position	...
Filter 2	-	Direction: POS = 1, NEG = 0	1
Sample volume (µL)	8	Check for NEG KIN/IR	0
Reagent 1 volume (µL)	400	Check for POS KIN/IR	1.000
Reagent 2 volume (µL)	-	Substrate Depletion	0.600
Reagent 1 position	...	Duplicate calibration	1
Reagent 2 position	-		
... Data entered by the operator		XXX Computed by the instrument	
* Assigned value			

Vers. 0505