TOTAL BILIRUBIN

Spectrophotometric
DIAZOTIZED SULFANILIC

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **AUTOLAB (=CORMAY LIVIA, SABA)**

Reagent preparation

Working Reagent: Transfer the contents of one Reagent B-T vial into a Reagent A-T bottle. Mix thoroughly.
Stable for 20 days at 2-8 ºC.

Instrument settings

BIL-T	BICHROMATIC	Standard value	*
Test	BIL-T	Standard position	...
Description	BIL-T	Factor	XXX
Unit	mg/dL	New factor memorization	1
Normal value minimum	0.00	Dilution ratio: 1:	2
maximum	1.10	Incubation time (sec)	120
Nº of decimals in result	2	Time for Reagent 2	
Model	0	T1	6
Type		T2	
Nº of blanks	1	Number of aspiration washes	2
Blank value minimum	0.000	Number of sampling washes	2
maximum	1.000	Volume of wash solution	500
Blank correction	1	Prediluted sample group	1
Linearity	15	Report position	...
Filter 1	540	Direction: POS = 1, NEG = 0	0
Filter 2	620	Duplicate calibration	1
Sample volume (µL)	40		
Reagent 1 volume (µL)	400		
Reagent 2 volume (µL)	-		
Reagent 1 position	...		
Reagent 2 position	-		
... Data entered by the operator		XXX Computed by the instrument	
* Assigned value			

Vers. 0207