

IRON

Spectrophotometric
CHROMAZUROL B

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **AIRONE 200 (=AUTOHUMALYZER 900 S)**

Reagent preparation

Reagent is ready to be used.

Instrument settings

UNITS	µg/dL
DECIMALS	0
REACTION TYPE	End Point
VOLUME SERUM	15
PLASMA	6
URINE	
REAGENT 1	300
REAGENT 2	0
ABS RANGE MIN.	0.0
ABS RANGE MAX.	2.000
LINEARITY LIMIT	500
REAGENT BLANK	Yes
CONTAMINATING	No
DIFERENTIAL	No
FILTER 1	620
FILTER 2	None
TIMES MIX 1	0.0
INCUBATION 1	240
LAG PHASE	2
MIX 2	0.0
INCUBATION 2	0
MEASURE	1
MEASURE TYPE	Std
FACTOR	0
NORMAL RANGE	55 – 155

Vers. 0212