

C-REACTIVE PROTEIN (CRP)

Turbidimetry
LATEX

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **AIRONE 200 (=AUTOHUMALYZER 900 S)**

Reagent preparation

Working Reagent: Pour the contents of a Latex vial into a Diluent bottle. Mix thoroughly.
Stable for 20 days at 2-8 °C.

Instrument settings

UNITS	mg/L
DECIMALS	1
REACTION TYPE	End Point
VOLUME SERUM	3
PLASMA	
URINE	
REAGENT 1	400
REAGENT 2	0
ABS RANGE MIN.	0.0
ABS RANGE MAX.	2.000
LINEARITY LIMIT	150
REAGENT BLANK	Yes
CONTAMINATING	No
DIFERENTIAL	No
FILTER 1	546
FILTER 2	None
TIMES MIX 1	0.2
INCUBATION 1	180
LAG PHASE	2
MIX 2	0.0
INCUBATION 2	0
MEASURE	1
MEASURE TYPE	Std
FACTOR	0.0
NORMAL RANGE	0.0 – 5.0

Vers. 0409