

TRANSFERRIN

Turbidimetry

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **AIRONE 200 (=AUTOHUMALYZER 900 S)**

Reagent preparation

Reagent is ready to use.

Instrument settings

UNITS	mg/dL
DECIMALS	0
REACTION TYPE	End Point
VOLUME SERUM	3
PLASMA	
URINE	
REAGENT 1	400
REAGENT 2	
ABS RANGE MIN.	0.0
ABS RANGE MAX.	2.000
LINEARITY LIMIT	700
REAGENT BLANK	Yes
CONTAMINATING	Yes
DIFERENTIAL	Yes
FILTER 1	546
FILTER 2	None
TIMES MIX 1	0.0
INCUBATION 1	300
LAG PHASE	2
MIX 2	0.0
INCUBATION 2	0
MEASURE	1
MEASURE TYPE	Std
FACTOR	00
NORMAL RANGE	200 - 360

Vers. 0409