

COMPLEMENT COMPONENT C4

Turbidimetry

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **CLINLINE 150 (=LISA, MASCOTT Plus)**

Reagent preparation

Reagent is ready to use.

Instrument settings

TEST NAME	C4	BLANK = STD	YES
SHORT NAME	C 4	N° OF STD	5
UNITS	mg/dL	STANDARD 1 VALUE	*
ASSAY TYPE	E.P. STD	STANDARD 1 POSITION	...
FILTER VALUE	340	STANDARD 2 VALUE	*
1 ST READING = 0	NO	STANDARD 2 POSITION	...
LAG PHASE 1	0	N° REP. STANDARD/CONTROL	1
N° OF MEASURES	10	CONTROL VALUE	*
REAGENT 1 VOLUME	280	CONTROL POSITION	...
REAGENT 1 DILUTION	0	CONTROL DEVIATION	*
REAGENT 1 POSITION	...	PREDILUTION RATE	1
REAGENT 2 VOLUME		POSTDILUTION RATE	5
REAGENT 2 DILUTION		DILUENT	WATER
REAGENT 2 POSITION		RINSE TYPE	3
SAMPLE VOLUME	7	UP. NORMAL VALUE	40
SAMPLE DILUENT		LOW NORMAL VALUE	10
ACTIVATION	ANY	LINEARITY LIMIT	90
LAG PHASE 2T		LOWER BLANK LIMIT	0
FACTOR		UPPER BLANK LIMIT	1500
STD CALCULATION	4 DEG		
... Data entered by the operator			
* Assigned value			

Version 0505