

DIRECT BILIRUBIN

Spectrophotometric
DIAZOTIZED SULFANILIC

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instruments: COBAS MIRA / MIRA Plus

Reagent preparation

Reagent 1: Use the Reagent A-D.

Reagent 2: Use the Reagent B-D.

Instrument settings

GENERAL			CALCULATION		
MEASUREMENT MODE	ABSORB		SAMPLE LIMIT	NO	SPACE
REACTION MODE	R- S-SR1	3	REACTION DIRECTION	INCREASE	1
CALIBRATION MODE	CALIBRATOR	2	CHECK	OFF	
REAGENT BLANK	REAG/DIL	2	CONVERSION FACTOR		1
CLEANER	NO	1	OFFSET		0
WAVELENGTH	550 nm	4	TEST RANGE LOW		0
DECIMAL POSITION		2 (mg/dL)	HIGH		15 mg/dL
UNIT		12 (mg/dL)	NORM RANGE LOW		0.0 mg/dL
			HIGH		0.25 mg/dL
ANALYSIS			NUMBER OF STEPS		1
POST DIL. FACTOR	NO	SPACE	CALCULATION STEP A	ENDPOINT	1
CONC. FACTOR	NO	SPACE	READING FIRST		4
SAMPLE CYCLE		2	LAST		17
VOLUME		20 µL			
DILUTION NAME	H2O	0	CALIBRATION		
VOLUME		10 µL	CALIB. INTERVAL	ON REQUEST	3
REAGENT CYCLE		1	TIME	NO	SPACE
VOLUME		160 µL	BLANK RANGE LOW	NO	SPACE
START R1 CYCLE		5	HIGH	NO	SPACE
VOLUME		40 µL	REAGENT RANGE LOW	NO	SPACE
DILUTION NAME	H2O	0	HIGH	NO	SPACE
VOLUME		20 µL			
STANDARD POS.	(*)				
STD-1	(**)				
(*) Enter the position of the CAL rack					
(**) Enter the value of the calibrator					