

CREATININE

Kinetic-spectrophotometric
ALKALINE PICRATE

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: HITACHI 704

Reagent preparation

Working Reagent: Mix equal volumes of Reagent A and Reagent B. Mix thoroughly.
Stable for 2 months at 2-8 °C.

Instrument settings

TEST	CREA
ASSAY CODE	RATE A - 3 - 10
SAMPLE VOLUME (µL)	20
R1 VOLUME (µL)	300 - 20 - 0
R2 VOLUME (µL)	0 - 20 - 0
WAVELENGTH (nm)	700 - 505
CALIB. METHOD	1 - 0 - 0
STD 1 CONC. POS.	0.00 - 1
STD 2 CONC. POS.	(*) - 2
STD 3 CONC. POS.	0 - 0
STD 4 CONC. POS.	0 - 0
STD 5 CONC. POS.	0 - 0
STD 6 CONC. POS.	0 - 0
UNIT	mg/dL
SD LIMIT	0.1
DUPLICATE LIMIT	100
SENSITIVITY LIMIT	0
ABS. LIMIT	4500 - 0
PROZONE LIMIT	0 - 0
EXPECTED VALUE	0.5 - 1.1
INSTRUMENT FACTOR	1.0
(*) Input the assigned value of the calibrator	

Vers. 0203