

IMMUNOGLOBULIN A (IgA)

Turbidimetry

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **HITACHI 902 (=HITACHI 7020)**

Reagent preparation

Reagent is ready to use.

Instrument settings

1 Test Name	IG A	29 Calib. Conc. 4	(*)
2 Assay Code (Mthd)	1 Point	30 Calib. Position 4	...
3 Assay Code (2. Test)	0	31 Calib. Conc. 5	(*)
4 Reaction Time	10	32 Calib. Position 5	...
5 Assay Point 1	35	33 Calib. Conc. 6	(*)
6 Assay Point 2	0	34 Calib. Position 6	...
7 Assay Point 3	0	35 S1 ABS	
8 Assay Point 4	0	36 K Factor	
9 Wavelength (SUB)	800	37 K2 Factor	0
10 Wavelength (MAIN)	340	38 K3 Factor	0
11 Sample Volume	3	39 K4 Factor	0
12 R1 Volume	300	40 K5 Factor	0
13 R1 Position	...	41 A Factor	0
14 R1 Bottle size	Small	42 B Factor	0
15 R2 Volume	0	43 C Factor	0
16 R2 Position	0	44 SD Limit	300
17 R2 Bottle size	Small	45 Duplicate Limit	400
18 R3 Volume	0	46 Sensitivity Limit	0
19 R3 Position	0	47 S1 ABS. Limit (L)	-32000
20 R3 Bottle size	Small	48 S1 ABS. Limit (H)	32000
21 Calib. Type (Type)	Logit-Log 4 (P)	49 ABS. Limit	0
22 Calib. Type (Weight)	0	50 ABS. Limit (D/I)	Increase
23 Calib. Conc. 1	0	51 Prozone Limit	0
24 Calib. Position 1	...	52 Prozone Limit (U/D)	Lower limit
25 Calib. Conc. 2	(*)	53 Prozone (End Point)	35
26 Calib. Position 2	...	54 Expected Value (L)	70
27 Calib. Conc. 3	(*)	55 Expected Value (H)	400
28 Calib. Position 3	...	56 Instrument Factor (a)	1.0
...	...	57 Instrument Factor (b)	0.0
... Data entered by the operator		58 Key Setting	...
* Assigned values			