

# PROTEIN

Spectrophotometric  
BIURET

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **ILAB 900 (=SHIMADZU CL-7200/7300)**

## Reagent preparation

Reagent is ready to be used.

## Instrument settings

<b>TEST N°:</b>		<b>CURVE N°:</b>	
TEST NAME	PROTEINS	NORMAL RANGE MALE L - H	65 – 80
BAR CODE		FEMALE L - H	65 – 80
MEASURING METHOD	Endpoint	VALID RANGE LL - HH	0 – 150
CALIB. METHOD	1 point	CORRECTION CONST. 1,2	1 – 0
UNIT	g/L	W1 – ABS INC / DEC, LIMIT	Inc – 2000
MEASURING POINT	43	HEMOLYSIS LIMIT	-
PHOTOMETRIC METHOD	Double	ICTERIC LIMIT	-
WAVELENGTHS 1, 2, 3	540 – 700	LIPEMIC LIMIT	-
SAMPLE VOL. N, D, I	6 – 3 – 12	CALIB. REPLICATES	5
R1 V. – STIR MODE (Nor)	300 – Yes – 0	REPLICATE, SENS. LIMITS	-
R2 V. – STIR MODE	-	SLOPE – OFFSET LIMITS	-
N°. DECIMAL POINTS	0	PARTNER TEST, CONST.	-
SAMPLE TYPE	Serum	PROZONE INC / DEC, LIMIT	-
... Data entered by the operator		* assigned value	