

TRANSFERRIN

Turbidimetry

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: **ILAB 900 (=SHIMADZU CL-7200/7300))**

Reagent preparation

Reagent is ready to use.

Instrument settings

TEST N°:		CURVE N°:	
TEST NAME		NORMAL RANGE MALE L - H	200 - 360
	TRANS		
BAR CODE		FEMALE L - H	200 - 360
MEASURING METHOD	Endpoint	VALID RANGE LL - HH	4.8 - 700
CALIB. METHOD	1 point	CORRECTION CONST. 1,2	1 - 0
UNIT	mg/dL	W1 - ABS INC / DEC, LIMIT	Inc - 3000
MEASURING POINT	25	HEMOLYSIS LIMIT	-
PHOTOMETRIC METHOD	Single	ICTERIC LIMIT	-
WAVELENGTHS 1, 2, 3	546	LIPEMIC LIMIT	-
SAMPLE VOL. N, D, I	3 - 2 - 30	CALIB. REPLICATES	5
R1 V. - STIR MODE (Nor)	300 - Yes - 0	REPLICATE, SENS. LIMITS	2500 - 2500
R2 V. - STIR MODE (Nor)	0 - Yes - 0	SLOPE - OFFSET LIMITS	999 - 2500
N°. DECIMAL POINTS	0	PARTNER TEST, CONST.	-
SAMPLE TYPE	Serum	PROZONE INC / DEC, LIMIT	-
Multipoint calibration	6 standards	... Data entered by the operator	
Function	5P-log-logit	* assigned value	

Version 0410