

DIRECT BILIRUBIN

Spectrophotometric
DIAZOTIZED SULFANILIC

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instrument: RA-XT (=OPERA)

Reagent preparation

Working Reagent: Transfer the contents of one Reagent B-D vial into a Reagent A-D bottle. Mix thoroughly.
Stable for 20 days at 2-8 °C.

Instrument settings

CHEM. #	(*)
NAME	BIL-D
IMMUNOASSAY	NO
TYPE	ENDPOINT
INVERSE	NO
% SAMPLE VOL.	60 (30 µL)
FILTER P.	550
BIC. CHEM.	NO
DELAY TIME	2:00
BLANK TYPE	DEFAULT
% REAGENT VOL.	60 (300 µL)
2 ND REAGENT	YES
% 2 ND REAGENT	15 (75 µL)
A2 DELAY	2:00
UNITS	mg/dL
UNIT FACTOR	1.000
DECIMAL	2
RBL. LOW	0.000
RBL. HIGH	0.040
RANGE LOW	0.00
RANGE HIGH	15.0
CAL. FACTOR	(**)
STANDARD VALUE	(***)
REAGENT RATE	0.000
NORMAL LOW	0.0
NORMAL HIGH	0.25
SLOPE	1.00
INTERCEPTION	0.00
ENDPOINT LIMIT	0.050
	(*) Data entered by the operator
	(**) CAL. FACTOR is determined by a calibration assay
	(***) Assigned value of the standard