

# GAMMA-GLUTAMYLTRANSFERASE

Continuous-spectrophotometric  
IFCC

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

## Instruments: **BT TARGA 2000/3000 (=ATAC 8000)**

### Reagent preparation

Working Reagent: Pour the contents of the Reagent B into the Reagent A bottle. Mix gently.  
Stable for 2 months at 2-8 °C.

### Instrument settings

Test Type		Kinetic Single
Serum Starter		Inactive
Filters	(A/B)	405 / 700
Units		U/L
Test Method		With factor
Test Methodology		
Number of Washes		1
Delay Time	(Sec)	0
Inc. Reagent	(Sec)	60
Reading Time	(Sec)	90
Test Limit	(Conc)	300
Max Abs Delta	(mABS)	250
Reagent mABS Limit		1100
Reagent A/B	(µL)	400
Reaction Direction		Increasing
Reagent Dilution		1:1
Initial ABS	(mABS)	1300
Curve Acceptance		100 %
Automatic Profile	(Inactive)	
Rerun Test Rgt BLK H:M		00:00
<b>Serum Parameters</b>		
Test's Name		Gamma-glutamyltransferase
Sample Volume	(µL)	40
Dilution Ratio		1:4
Min. Max. M		15 – 86
Min. Max. F		10 – 40
Min. Max. B		
(*) Input the assigned value of the calibrator		