

LACTATE DEHYDROGENASE

Continuous-spectrophotometric
PYRUVATE

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instruments: **BT TARGA 2000/3000 (=ATAC 8000)**

Reagent preparation

Working Reagent: Pour the contents of the Reagent B into the Reagent A bottle. Mix gently.
Stable for 2 months at 2-8°C.

Instrument settings

Test Type		Kinetic Single
Serum Starter		Inactive
Filters	(A/B)	340 / 700
Units		U/L
Test Method		With factor
Test Methodology		
Number of Washes		1
Delay Time	(Sec)	0
Inc. Reagent	(Sec)	90
Reading Time	(Sec)	120
Test Limit	(Conc)	1200
Max Abs Delta	(mABS)	200
Reagent mABS Limit		800
Reagent A/B	(µL)	400
Reaction Direction		Decreasing
Reagent Dilution		1:1
Initial ABS	(mABS)	1999
Curve Acceptance		100 %
Automatic Profile	(Inactive)	
Rerun Test Rgt BLK H:M		00:00
Serum Parameters		
Test's Name		Lactate dehydrogenase
Sample Volume	(µL)	8
Dilution Ratio		1:2
Min. Max. M		207 – 414
Min. Max. F		207 – 414
Min. Max. B		
(*) Input the assigned value of the calibrator		