

LIPASE

Color

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instruments: **BT TARGA 2000/3000 (=ATAC 8000)**

Reagent preparation

Reagent 1: use the Reagent A.
 Reagent 2: use the Reagent B.

Instrument settings

Test Type		Kinetic Starter
Serum Starter		Inactive
Filters	(A/B)	570 / 700
Units		U/L
Test Method		With factor
Test Methodology		
Number of Washes		1 2
Delay Time	(Sec)	180
Inc. Reagent	(Sec)	300 180
Reading Time	(Sec)	60
Test Limit	(Conc)	400
Max Abs Delta	(mABS)	200
Reagent mABS Limit		1000
Reagent A/B	(μ L)	200 100
Reaction Direction		Increasing
Reagent Dilution		1:1
Initial ABS	(mABS)	1999
Curve Acceptance		100 %
Automatic Profile	(Inactive)	
Rerun Test Rgt BLK H:M		00:00
Serum Parameters		
Test's Name		Lipase
Sample Volume	(μ L)	3
Dilution Ratio		1:2
Min. Max. M		12 - 48
Min. Max. F		12 - 48
Min. Max. B		
(*) Input the assigned value of the calibrator		