

APOLIPOPROTEIN A1 (Apo A1)

Turbidimetry

The application parameters comprised here constitute a guide to facilitate the validation of our reagents by the instrument. It is advisable to validate the use when there is any change in software or reagent versions.

Instruments: **BT TARGA 2000/3000 (=ATAC 8000)**

Reagent preparation

Reagent 1: Use the Reagent A.

Reagent 2: Use the Reagent B.

Instrument settings

Test Type		S.B. (A) with Starter
Serum Starter		Inactive
Filters	(A/B)	340 / 700
Units		mg/dL
Test Method		2-Logit 5
Test Methodology		Immunoturb
Number of Washes		1 2
Delay Time	(Sec)	0
Inc. Reagent	(Sec)	60 0
Reading Time	(Sec)	300
Test Limit	(Conc)	250
Max Abs Delta	(mABS)	0.00
Reagent mABS Limit		1900
Reagent A/B	(µL)	360 90
Reaction Direction		Increasing
Reagent Dilution		1:1
Initial ABS	(mABS)	0.00
Curve Acceptance		100 %
Automatic Profile	(Inactive)	Inactive
Rerun Test Rgt BLK H:M		00:00
Serum Parameters		
Test's Name		Apolipoprotein A1
Sample Volume	(µL)	3
Dilution Ratio		1:1
Min. Max. M		94 - 178
Min. Max. F		94 - 178
Min. Max. B		
(*) Input the assigned value of the calibrator		